



Clinical Quality MRI Image under Inhomogeneous Magnetic Field

IP Status: Issued US Patent; **Application #:** 16/341,677

Clinical quality images

This technology is a new MRI (Magnetic Resonance Imaging) acquisition sequence implemented in software that provides clinical quality images under extremely inhomogeneous magnetic fields. The novel method employs radiofrequency (RF) pulse trains for signal excitation and acquires refocusing echoes generated by the RF pulse train excitations. It does not use refocusing RF pulses.

This new steady state sequence uses a train of low-flip-angle (e.g., 5°) short duration (e.g., tens to hundreds of microseconds) pulses and a wide bandwidth for signal excitation. It is specifically designed to yield robust data in the face of extreme B₀ inhomogeneity while limiting RF peak power and specific absorption rate (SAR).

A modified version of this MP-SSFP sequence can also be used to map the three-dimensional (3D) distribution of B₀, even when the degree of B₀ inhomogeneity is extremely large. For example, high quality three-dimensional (3D) images and B₀ mapping have been demonstrated using a magnet with 200 kHz of resonance frequency variation across the object of interest. In a clinical 1.5 Tesla magnet, this degree of B₀ inhomogeneity corresponds to ~3000 ppm.

Overcomes limitations of inhomogeneous fields

Magnetic resonance imaging (MRI) under highly inhomogeneous static magnetic fields faces fundamental difficulties in signal excitation and data acquisition. The conventional approach of using spin-echo sequences is limited by the requirement to use short, high-flip-angle (180°) refocusing pulses which results in high SAR in subjects. This new approach introduces an MRI method under inhomogeneous magnetic fields that uses only short, low-flip-angle RF pulses. A 3D B₀ map can be measured by using phase-encoding in all three spatial directions.

Phase of Development

- Prototype developed. Validation in humans using low-cost magnet.

Benefits

- Clinical quality images with low cost, extremely inhomogeneous magnets

Features

- Image-based method for 3D mapping of the MRI magnetic field
- Low flip angle pulses provide large bandwidth while maintaining low RF power and SAR
- Complements other MRI contrasts (eg, T1- and T2*-weighted MRI) by providing T2-weighted images and/or diffusion-weighted images

Applications

Technology ID

20170013

Category

Engineering & Physical Sciences/MRI & Spectroscopy
Life Sciences/Diagnostics & Imaging
Life Sciences/MRI & Spectroscopy
Software & IT/Algorithms

Learn more



- Magnetic resonance imaging (MRI)
- Portable MRI scanners
- Image-based mapping of the magnetic field of MRI magnets

Researchers

Michael Garwood, PhD

Professor, Department of Radiology, Center for Magnetic Resonance Research (CMRR)

[External Link](http://www.radiology.umn.edu) (www.radiology.umn.edu)

Naoharu Kobayashi, PhD

Assistant Professor, Department of Radiology, Center for Magnetic Resonance Research (CMRR)

[External Link](http://www.radiology.umn.edu) (www.radiology.umn.edu)

Djauat Idiyatullin, PhD

Assistant Professor, Department of Radiology, Center for Magnetic Resonance Research (CMRR)

[External Link](http://www.radiology.umn.edu) (www.radiology.umn.edu)

Interested in Licensing?

The University relies on industry partners to further develop and ultimately commercialize this technology. The license is for the sale, manufacture or use of products claimed by the patents. Please contact us to share your business needs and licensing and technical interests in this technology.