

Designing 2D and 3D Adiabatic Pulses

IP Status: Issued US Patent; Application #: 15/493,959

Multi-Dimensional Adiabatic Pulse Design Using a Sub-Pulse Approach

A new method designs and implements spatially selective, multidimensional adiabatic radio frequency (RF) pulses for use in magnetic resonance imaging (MRI). Spatially selective inversion is achieved adiabatically in both two- and three-dimensional (2D and 3D) regions of interest. The multidimensional adiabatic pulses are designed by dividing a parent adiabatic pulse into 2D selective sub-pulse elements. Selective excitation is achieved by the sub-pulses while the phases and amplitudes of the sequence of sub-pulses are prescribed according to an adiabatic full passage (the parent adiabatic pulse). The approach can be extended to 3D by applying blips along the remaining direction between sub-pulses.

Complete Spatiotemporal Properties

Previous methods for designing a 2D adiabatic pulse use 1D amplitude-modulated sub-pulses to define a slice in one spatial direction and frequency- and amplitude-modulation of these consecutive sub-pulses to define a slice in a second spatial direction. By using 2D sub-pulses, this new approach enables 3D selection in space. By exploiting the spatiotemporal properties of this 3D excitation, it is possible to perform 3D spatiotemporally-encoded MRI (rather than exciting the entire object simultaneously). Furthermore, it offers the flexibility to choose the desired modulation method in either direction and/or to choose the desired shape of the selective excitation (square, cylinder, etc. Other advantages include significant acceleration in STEREO (scan-time), better image fidelity, compensation for B0 inhomogeneities in 1D, and volumetric slab selection for SWIFT.

BENEFITS AND FEATURES:

- Spatially selective, multidimensional adiabatic radio frequency (RF) pulses
- 2D and 3D adiabatic pulse design
- Complete spatiotemporal properties
- Uses a sub-pulse approach
- Significant acceleration in STEREO (scan-time)
- Image fidelity better k-space trajectory to reduce saturation in center of k-space
- Can compensate for B0 inhomogeneities in 1D
- Provides volumetric slab selection for SWIFT (faster, higher resolution)
- Software implemented on MRI scanner
- Potentially disruptive (low-cost) technology

APPLICATIONS:

- May enable anatomy specific (breast, brain) scanners
- Could expand customer base beyond hospitals and imaging centers
- Selectively exciting a desired shape adiabatically can be used in applications such as navigator, in which the pulse can be tailored to a target organ to track anatomic motion.
- Can be used to selectively excite a volume of interest (e.g., the heart) to permit high resolution imaging in the volume in reduced time (i.e., since the field-of-view is reduced by volume selection, the number of k-space samples required to achieve a given resolution is reduced).

Technology ID

20160260

Category

Engineering & Physical Sciences/MRI & Spectroscopy Life Sciences/Diagnostics & Imaging Life Sciences/Human Health Life Sciences/Medical Devices Life Sciences/MRI & Spectroscopy Software & IT/Algorithms

View online page



Phase of Development - Pilot scale demonstration

Researchers

Michael Garwood, PhD Professor, Department of Radiology, Center for Magnetic Research (CMRR) External Link (www.med.umn.edu) Albert Jang, PhD

Publications

Designing 3D selective adiabatic radiofrequency pulses with single and parallel transmission Magnetic Resonance in Medicine, 2017 May 12